

Proficiency testing for in-house and external measuring stations - results and evaluation

Proficiency testing scheme: Inorganic acids with own sampling

20 – 21 March 2018

Part 1: Volatile inorganic acids

Summary of laboratory test results

Measurand hydrochloric acid

Laboratory	Sample 1	Z score	Sample 2	Z score	Sample 3	Z score
Unit	mg/m ³		mg/m ³		mg/m ³	
3	0.460	-1.56	4.450	-0.92	2.690	-2.20 E
5	0.518	-0.50	4.622	-0.57	3.247	-0.59
6	0.625	1.47	5.569	1.37	4.233	2.27 E
64	1.200	12.01 BE	5.000	0.21	3.700	0.73
66	0.560	0.27	4.670	-0.47	3.440	-0.03
111	0.590	0.82	5.077	0.36	3.510	0.18
130	0.508	-0.68	4.750	-0.30	3.100	-1.01
212	0.538	-0.13	4.710	-0.39	3.250	-0.58
243	0.547	0.04	4.832	-0.14	3.319	-0.38
259	0.560	0.27	5.310	0.84	4.000	1.60
263	0.195	-6.42 BE	3.150	-3.57 BE	0.132	-9.62 BE
-	-	--	-	--	-	--
Method	ISO 5725-2		ISO 5725-2		ISO 5725-2	
Assessment	Z <=2.00		Z <=2.00		Z <=2.00	
No. of laboratories that submitted results	11		11		11	
Mean	0.545		4.899		3.449	
Reproducibility s.d.	0.048		0.342		0.445	
Rel. reproducibility s.d.	8.77 %		6.98 %		12.90 %	
Reference value	0.590		4.810		3.340	
Target s.d.	0.055		0.490		0.345	
Rel. target s.d.	10.00 %		10.00 %		10.00 %	
Lower limit of tolerance	0.436		3.919		2.759	
Upper limit of tolerance	0.654		5.879		4.139	
Type B outliers	2		1		1	
No. of measurement values outside of tolerance limits	2		1		3	

Laboratory	Sample 1 Z score	Sample 2 Z score	Sample 3 Z score
Explanation of outlier types			
A: Single outlier	Grubbs		
B: Differing laboratory mean	Grubbs		
C: Excessive laboratory s.d.	Cochran		
E: mean outside tolerance limits			
F: Z-Score >3.5			
L: Differing laboratory mean (Grubbs II)	Grubbs für 2		

Summary of laboratory test results


Measurand nitric acid

Laboratory	Sample 1	Z score	Sample 2	Z score	Sample 3	Z score
Unit	mg/m ³		mg/m ³		mg/m ³	
3	1.850	0.15	2.510	0.25	2.210	-0.86
5	1.449	-2.05 E	2.261	-0.77	2.216	-0.84
6	1.865	0.23	2.704	1.04	2.701	1.17
64	1.900	0.42	2.500	0.21	2.500	0.34
66	2.170	1.90	2.580	0.53	2.590	0.71
111	1.875	0.28	2.451	0.01	2.420	0.01
130	1.660	-0.90	2.250	-0.81	2.250	-0.69
212	1.758	-0.36	2.180	-1.10	2.310	-0.44
243	1.915	0.50	2.435	-0.06	2.561	0.59
259	1.790	-0.18	2.620	0.70	3.200	3.24 BE
263	0.154	-9.16 BE	0.408	-8.33 BE	0.086	-9.64 BE
-	-	--	-	--	-	--
Method	ISO 5725-2		ISO 5725-2		ISO 5725-2	
Assessment	Z <=2.00		Z <=2.00		Z <=2.00	
No. of laboratories that submitted results	11		11		11	
Mean	1.823		2.449		2.417	
Reproducibility s.d.	0.186		0.172		0.181	
Rel. reproducibility s.d.	10.20 %		7.01 %		7.47 %	
Reference value	1.890		2.500		2.820	
Target s.d.	0.182		0.245		0.242	
Rel. target s.d.	10.00 %		10.00 %		10.00 %	
Lower limit of tolerance	1.459		1.959		1.934	
Upper limit of tolerance	2.188		2.939		2.901	
Type B outliers	1		1		2	
No. of measurement values outside of tolerance limits	2		1		2	

Laboratory	Sample 1 Z score	Sample 2 Z score	Sample 3 Z score
Explanation of outlier types			
A: Single outlier	Grubbs		
B: Differing laboratory mean	Grubbs		
C: Excessive laboratory s.d.	Cochran		
E: mean outside tolerance limits			
F: Z-Score >3.5			
L: Differing laboratory mean (Grubbs II)	Grubbs für 2		


Summary results

Sample:	Sample 1	Mean:	0.545 mg/m ³
Measurand:	hydrochloric acid	Reproducibility s.d.:	0.048 mg/m ³
Method:	ISO 5725-2	Rel. reproducibility s.d.:	8.77%
Rel. target s.d.:	10.00% (Limited)	Reference value:	0.590 mg/m ³
No. of laboratories:	9	Range of tolerance:	0.436 - 0.654 mg/m ³ (Z-Score <= 2.00)




Summary results

Sample:	Sample 2	Mean:	4.899 mg/m ³
Measurand:	hydrochloric acid	Reproducibility s.d.:	0.342 mg/m ³
Method:	ISO 5725-2	Rel. reproducibility s.d.:	6.98%
Rel. target s.d.:	10.00% (Limited)	Reference value:	4.810 mg/m ³
No. of laboratories:	10	Range of tolerance:	3.919 - 5.879 mg/m ³ (Z-Score <= 2.00)




Summary results

Sample:	Sample 3	Mean:	3.449 mg/m ³
Measurand:	hydrochloric acid	Reproducibility s.d.:	0.445 mg/m ³
Method:	ISO 5725-2	Rel. reproducibility s.d.:	12.90%
Rel. target s.d.:	10.00% (Limited)	Reference value:	3.340 mg/m ³
No. of laboratories:	10	Range of tolerance:	2.759 - 4.139 mg/m ³ (Z-Score <= 2.00)




Analyte chart of Z-Scores

Measurand: hydrochloric acid




Summary results

Sample:	Sample 1	Mean:	1.823 mg/m ³
Measurand:	nitric acid	Reproducibility s.d.:	0.186 mg/m ³
Method:	ISO 5725-2	Rel. reproducibility s.d.:	10.20%
Rel. target s.d.:	10.00% (Limited)	Reference value:	1.890 mg/m ³
No. of laboratories:	10	Range of tolerance:	1.459 - 2.188 mg/m ³ (Z-Score <= 2.00)




Summary results

Sample:	Sample 2	Mean:	2.449 mg/m ³
Measurand:	nitric acid	Reproducibility s.d.:	0.172 mg/m ³
Method:	ISO 5725-2	Rel. reproducibility s.d.:	7.01%
Rel. target s.d.:	10.00% (Limited)	Reference value:	2.500 mg/m ³
No. of laboratories:	10	Range of tolerance:	1.959 - 2.939 mg/m ³ (Z-Score <= 2.00)




Summary results

Sample:	Sample 3	Mean:	2.417 mg/m ³
Measurand:	nitric acid	Reproducibility s.d.:	0.181 mg/m ³
Method:	ISO 5725-2	Rel. reproducibility s.d.:	7.47%
Rel. target s.d.:	10.00% (Limited)	Reference value:	2.820 mg/m ³
No. of laboratories:	9	Range of tolerance:	1.934 - 2.901 mg/m ³ (Z-Score <= 2.00)



Analyte chart of Z-Scores

Measurand: nitric acid



Questions and Answers

Participant	Sample carrier HCl/HNO ₃	Prefilter for chlorid/nitrat
3	37 mm Quarzfaserpalnfilter (Munktell), imprägniert mit Natriumcarbonatlösung	ja
5	Imprägnierte Glasfaserfilter	Ja
6	imprägnierter Quarzfaserfilter	ja: nicht imprägnierter Quarzfaserfilter
64	Quarzfaserfilter M37	nein
66	Quarzfaserfilter	ja
111	Quarzfaserfilter, imprägniert mit 1 mol/L Na ₂ CO ₃	Ja, nicht imprägnierter Quarzfaserfilter
130	37 mm Quarzplanfilter Munktell, mit 1,0 mol/L Natriumcarbonatlg.	nein
212	Quarzfaserfilter imprägniert mit Na ₂ CO ₃	Ja
243	Impinger (VE-Wasser)	nein
259	imprägnierte Quarzfaserfilter	ja
263	Quarzfaserfilter 37mm imprägniert mit 1 mol/l Na ₂ CO ₃	Ja

Participant	Sampling pump	Volume flow
3	A110-E, Fa. Honold GmbH & Co KG, Nümbrecht,	2 l/min
5	Gilian GilAir	2000 ml/min
6	GSA 4000ex	2 L/min
64	GSA SG 5100	2l
66	SKC Deluxe (ATEX)	2 L/min
111	GilairPlus	2 L/min
130	Model PCMTX8. SKC Inc	2,0 L/min, bzw . 1,0 L/min
212	GSA 5100ex, SKC	2 L/min; 1 L/min (test gas 2)
243	GilAir Plus	1,330 L/min (80 L/h)
259	Gilian 5000	2,0 l/min

Participant	Volume flow measurement
3	Pumpe mit elektronischer Volumenstromregelung auf Basis Massflow meter, Pumpenfluss wird jährlich mit Referenzgasuhr überprüft und mit Korrekturfaktor versehen

Volatile inorganic acids 2/2018

Participant	Volume flow measurement
-------------	-------------------------

5	Defender 530
6	TSI 4100
66	Analyt-MTC Serie 358MLW
111	Bios DC Lite
130	Gilibrator
212	TSI 4146
243	Drycal Defender DC-Lite
259	Rotameter

Participant	Sampling time	Analytical method
3	140 min bzw . 15 min (MR3)	BGIA Arbeitsmappe 6172 bzw . 6173
5	2 h oder 0,5 h	IFA 6172, IFA 6173
6	2 h bzw . 15 min	IFA-Arbeitsmappe 6172 bzw . 6173
64	2,0 bzw 0,25	IFA Arbeitsmappe Nr. 6172 und 6173
66	120 min bzw . 15 min	IFA 6172
111	120 min, bzw . 15 Minuten für den Kurzzeitwert	IFA 6172, bzw . IFA 6173
130	Run 1 + 2: je 3 x 2 h, Run 4: 2 x 15 min + 2 x 15 min	IFA-Arbeitsmappe 6172
212	120; 15 (test gas 3)	IFA 6172
243	120 min (Probe 1 und 2) bzw . 15 min (Probe 3)	HCl/HNO3: validierte Hausmethode; H2SO4/H3PO4: IFA 6173
259	120 min, 60 min, 15 min	IFA 6172

Participant	Desorption solution
-------------	---------------------

3	Reinstw assen
5	VE-Wasser
6	Reinstw assen
64	entionisiertes Wasser
66	Wasser
111	dest. Wasser

Volatile inorganic acids 2/2018

Participant	Desorption solution
-------------	---------------------

130	Reinstw assen
212	Reinstw assen
243	HCL/HNO3: nicht zutreffend (Probenahme im Impinger); H2SO4/H3PO4: vom Ringversuchsveranstalter überschichtet
259	Reinstw assen
263	H2O nanopur

Participant	Volume of desorption solution
-------------	-------------------------------

3	10 mL
5	10 ml und 8 ml
6	10 mL
64	10 ml
66	10 mL
111	5ml / 10ml
130	10 mL
212	10 ml
243	HCL/HNO3: nicht zutreffend (Probenahme im Impinger); H2SO4/H3PO4: vom Ringversuchsveranstalter überschichtet
259	10 ml
263	10 ml

Participant	Desorption time
-------------	-----------------

3	15 min Ultraschallbad und anschließend 30 min stehen lassen
5	15 min
6	15 min. Ultraschall, 30 min. stehen gelassen
64	15 min im Ultraschallbad, 30 min stehen lassen
66	ja
111	15 min Ultraschallbad
130	15 min US, dann 30 min stehengelassen
212	15 min US

Volatile inorganic acids 2/2018

Participant	Desorption time
-------------	-----------------

243	HCL/HNO3: nicht zutreffend (Probenahme im Impinger); H2SO4/H3PO4: Desorptionszeit unbekannt, 15 min Ultraschallbad
259	15 min im Ultraschallbad
263	1h Schüttelapparatur

Participant	Analytical column
-------------	-------------------

3	Anionensäule Metrosep A-SUPP 4 mm (ID)* 250 mm (Länge) (Metrohm Nr.: 6.1006.430)
6	AS19, 2x250mm
64	Ion Pac AS14A
66	Dionex ion pac AS 22 250*4 mm
111	IC 1: Trennsäule AS-22; IC 2: Trennsäule AS9-HC
130	AS22 4 mm von Dionex
212	Metrosep A Supp 4/5 Guard/ Metrosep A Supp 5-150
243	Vorsäule: Metrosep RP Guard (Metrohm); Trennsäule: Metrosep Anion Dual 2
259	AS 22
263	Metrosep A Supp 7 150/4.0

Participant	Ion Chromatographic System
-------------	----------------------------

3	IC Gerät, 861 Advanced, Firma Metrohm mit Leitfähigkeitsdetektor
6	Thermo ICS 5000
64	ICS 1100 Thermo Scientific; Leitfähigkeitsdetektor, ohne Autosampler
66	HP 1050 ASL, Dionex IC DX-120, Dionex Pulsed Electrochemical Detector
111	IC 1: Dionex ICS 1100, IC 2: Dionex ICS 2100; beide Leitfähigkeitsdetektor; beide Autosampler: Dionex AS-AP
130	Dionex ICS 1100
212	Metrohm: 881 Compact IC pro / Leitfähigkeitsdetektor
243	IC mit Leitfähigkeitsdetektor Firma Metrohm
259	IC 1000 mit Autosampler
263	Metrohm Professional IC Vario, Leitfähigkeitsdetektor, AS Professional IC 850

Volatile inorganic acids 2/2018

Participant	Mobile phase	Flow rate
3	4 mmol/l NaHCO ₃ + 1 mmol/l Na ₂ CO ₃	1 mL/min
6	KOH	0,25 mL/min
64	8 mM Na ₂ CO ₃ , 1mM NaHCO ₃	1 ml/min
66	Natriumcarbonat/-hydrogencarbonat-Lösung	1,2
111	IC 1: 4,5 mmol/l Natriumcarbonat / 1,4 mmol/l Natriumhydrogencarbonat; IC 2: 9mmol/l Natriumcarbonat	IC 1: 0,3 ml/min; IC 2: 1 ml/min
130	4,5 mmol Na ₂ CO ₃ / 1,4 mmol NaHCO ₃	1,2 mL/min
212	Na ₂ CO ₃ /NaHCO ₃	0,7 mL/min
243	1,3 mmol Na ₂ CO ₃ / 2,0 mmol NaHCO ₃ mit 2 % Aceton	0,8 mL/min
259	NaHCO ₃ 1,4 mmmol/l, Na ₂ CO ₃ 4,5 mmol/l	1,2
263	3,6 mmol Na ₂ CO ₃	0,7

Participant	Recovery rate	Date of analysis
3	nein	04.04.2018
5		12.04.2018
6	nein	01.04.18
64	nein	11. bis 16. April 2018
66	nein	20.04.2018
111	nein	05.04.2018 - 16.04.2018
130	ja, Zielwertkarte	18.4.2018
212	Nein	9.4.2018; 11.4.2018
243	nein	22.03.2018 (HCl/HNO ₃); 06.04.2018 (H ₂ SO ₄ /H ₃ PO ₄)
259	Doppelbestimmung	11./12.04.2018
263	-	04.04.2018 / 24.04.2018

Proficiency testing for in-house and external measuring stations - results and evaluation

Proficiency testing scheme: Inorganic acids

March 2018

Part 2: Non-volatile inorganic acids

Summary of laboratory test results

Measurand phosphoric acid

Laboratory	Sample 1	Z score	Sample 2	Z score	Sample 3	Z score
Unit	mg/m ³		mg/m ³		mg/m ³	
3	0.296	-0.20	0.840	0.12	0.398	-0.24
5	0.319	0.56	0.867	0.44	0.429	0.51
6	0.310	0.26	0.867	0.45	0.471	1.55
10	0.288	-0.47	0.762	-0.82	0.357	-1.25
27	0.280	-0.73	0.610	-2.65 BE	0.350	-1.42
64	0.300	-0.07	0.800	-0.36	0.500	2.26 E
66	0.340	1.25	0.850	0.24	0.470	1.52
68	0.294	-0.27	0.885	0.66	0.458	1.23
72	0.312	0.33	0.850	0.24	0.422	0.35
73	0.305	0.11	0.805	-0.30	0.405	-0.08
74	0.290	-0.40	0.827	-0.03	0.384	-0.59
78	0.305	0.10	0.828	-0.02	0.410	0.05
83	0.282	-0.66	0.866	0.43	0.396	-0.28
95	0.327	0.82	0.860	0.36	0.430	0.54
110	0.310	0.26	0.820	-0.12	0.400	-0.19
111	0.312	0.33	0.834	0.05	0.415	0.17
114	0.321	0.63	0.826	-0.05	0.406	-0.05
130	0.300	-0.07	0.812	-0.22	0.407	-0.02
151	0.313	0.36	0.835	0.06	0.413	0.13
174	0.331	0.96	0.855	0.30	0.372	-0.88
177	0.301	-0.04	0.868	0.46	0.422	0.35
178	0.270	-1.06	0.800	-0.36	0.380	-0.68
195	0.200	-3.38 BE	0.750	-0.96	0.360	-1.17
208	0.313	0.36	0.806	-0.29	0.401	-0.17
224	0.285	-0.57	0.779	-0.61	0.372	-0.87
243	0.300	-0.07	0.845	0.18	0.436	0.69

Laboratory	Sample 1	Z score	Sample 2	Z score	Sample 3	Z score
248	0.309	0.23	0.859	0.35	0.426	0.44
259	0.290	-0.40	0.820	-0.12	0.390	-0.44
263	0.261	-1.37	0.702	-1.54 B	0.343	-1.58
266	0.297	-0.17	0.821	-0.11	0.413	0.13
–	–	–	–	–	–	–
Method	ISO 5725-2		ISO 5725-2		ISO 5725-2	
Assessment	Z ≤2.00		Z ≤2.00		Z ≤2.00	
No. of laboratories that submitted results	30		30		30	
Mean	0.302		0.830		0.408	
Reproducibility s.d.	0.018		0.033		0.036	
Rel. reproducibility s.d.	5.84 %		3.96 %		8.92 %	
Reference value	0.312		0.841		0.425	
Target s.d.	0.030		0.083		0.041	
Rel. target s.d.	10.00 %		10.00 %		10.00 %	
Lower limit of tolerance	0.242		0.664		0.326	
Upper limit of tolerance	0.363		0.996		0.489	
Type B outliers	1		2			
No. of measurement values outside of tolerance limits	1		1		1	
Explanation of outlier types						
A: Single outlier	Grubbs					
B: Differing laboratory mean	Grubbs					
C: Excessive laboratory s.d.	Cochran					
E: mean outside tolerance limits						
F: Z-Score >3.5						
L: Differing laboratory mean (Grubbs II)	Grubbs für 2					

Summary of laboratory test results


Measurand sulphuric acid

Laboratory	Sample 1	Z score	Sample 2	Z score	Sample 3	Z score
Unit	mg/m ³		mg/m ³		mg/m ³	
3	0.1290	-0.11	0.0630	-0.14	0.0900	-0.24
5	0.1361	0.43	0.0647	0.12	0.0956	0.36
6	0.1510	1.57	0.0830	2.99 E	0.1170	2.68 E
10	0.1250	-0.42	0.0600	-0.61	0.0838	-0.92
27	0.1300	-0.04	0.0800	2.52 E	0.1000	0.84
64	0.1400	0.73	0.0700	0.95	0.0900	-0.24
66	0.1300	-0.04	0.0690	0.80	0.0950	0.30
68	0.1160	-1.11	0.0630	-0.14	0.0860	-0.68
72	0.1420	0.88	0.0650	0.17	0.0980	0.62
73	0.1308	0.02	0.0643	0.06	0.0940	0.19
74	0.1430	0.96	0.0690	0.80	0.0980	0.62
78	0.1280	-0.19	0.0576	-0.99	0.0870	-0.57
83	0.0324	-7.52 BE	0.0161	-7.48 BE	0.0230	-7.51 BE
95	0.1310	0.04	0.0590	-0.77	0.0870	-0.57
110	0.1500	1.49	0.0900	4.08 FE	0.1100	1.93
111	0.1360	0.42	0.0680	0.64	0.0970	0.52
114	0.1359	0.41	0.0591	-0.75	0.0901	-0.23
130	0.1270	-0.27	0.0620	-0.30	0.0890	-0.35
151	0.1350	0.34	0.0630	-0.14	0.0920	-0.03
174	0.1360	0.42	0.0590	-0.77	0.0870	-0.57
177	0.1430	0.96	0.0690	0.80	0.1000	0.84
178	0.1000	-2.34 E	0.0500	-2.18 E	0.0800	-1.33
195	0.1000	-2.34 E	0.0600	-0.61	0.0800	-1.33
208	0.1270	-0.27	0.0619	-0.31	0.0880	-0.46
224	0.0941	-2.79 E	0.0426	-3.34 E	0.0656	-2.89 E
243	0.1310	0.04	0.0620	-0.30	0.0950	0.30

Laboratory	Sample 1	Z score	Sample 2	Z score	Sample 3	Z score
248	0.1360	0.42	0.0640	0.01	0.0980	0.62
259	0.1600	2.26 E	0.0800	2.52 E	0.1100	1.93
263	0.1176	-0.99	0.0601	-0.60	0.0822	-1.09
266	0.1240	-0.50	0.0612	-0.42	0.0897	-0.28
–	–	--	–	--	–	--
Method	ISO 5725-2		ISO 5725-2		ISO 5725-2	
Assessment	Z <=2.00		Z <=2.00		Z <=2.00	
No. of laboratories that submitted results	30		30		30	
Mean	0.1305		0.0639		0.0922	
Reproducibility s.d.	0.0147		0.0083		0.0101	
Rel. reproducibility s.d.	11.26 %		12.92 %		10.97 %	
Reference value	0.1340		0.0634		0.0937	
Target s.d.	0.0130		0.0064		0.0092	
Rel. target s.d.	10.00 %		10.00 %		10.00 %	
Lower limit of tolerance	0.1044		0.0511		0.0738	
Upper limit of tolerance	0.1566		0.0767		0.1107	
Type B outliers	1		1		1	
Type F outliers			1			
No. of measurement values outside of tolerance limits	5		7		3	
Explanation of outlier types						
A: Single outlier	Grubbs					
B: Differing laboratory mean	Grubbs					
C: Excessive laboratory s.d.	Cochran					
E: mean outside tolerance limits						
F: Z-Score >3.5						
L: Differing laboratory mean (Grubbs II)	Grubbs für 2					


Summary results

Sample:	Sample 1	Mean:	0.302 mg/m ³
Measurand:	phosphoric acid	Reproducibility s.d.:	0.018 mg/m ³
Method:	ISO 5725-2	Rel. reproducibility s.d.:	5.84%
Rel. target s.d.:	10.00% (Limited)	Reference value:	0.312 mg/m ³
No. of laboratories:	29	Range of tolerance:	0.242 - 0.363 mg/m ³ (Z-Score <= 2.00)




Summary results

Sample:	Sample 2	Mean:	0.830 mg/m ³
Measurand:	phosphoric acid	Reproducibility s.d.:	0.033 mg/m ³
Method:	ISO 5725-2	Rel. reproducibility s.d.:	3.96%
Rel. target s.d.:	10.00% (Limited)	Reference value:	0.841 mg/m ³
No. of laboratories:	28	Range of tolerance:	0.664 - 0.996 mg/m ³ (Z-Score <= 2.00)




Summary results

Sample:	Sample 3	Mean:	0.408 mg/m ³
Measurand:	phosphoric acid	Reproducibility s.d.:	0.036 mg/m ³
Method:	ISO 5725-2	Rel. reproducibility s.d.:	8.92%
Rel. target s.d.:	10.00% (Limited)	Reference value:	0.425 mg/m ³
No. of laboratories:	30	Range of tolerance:	0.326 - 0.489 mg/m ³ (Z-Score <= 2.00)




Analyte chart of Z-Scores

Measurand: phosphoric acid




Summary results

Sample:	Sample 1	Mean:	0.1305 mg/m ³
Measurand:	sulphuric acid	Reproducibility s.d.:	0.0147 mg/m ³
Method:	ISO 5725-2	Rel. reproducibility s.d.:	11.26%
Rel. target s.d.:	10.00% (Limited)	Reference value:	0.1340 mg/m ³
No. of laboratories:	29	Range of tolerance:	0.1044 - 0.1566 mg/m ³ (Z-Score <= 2.00)




Summary results

Sample:	Sample 2	Mean:	0.0639 mg/m ³
Measurand:	sulphuric acid	Reproducibility s.d.:	0.0083 mg/m ³
Method:	ISO 5725-2	Rel. reproducibility s.d.:	12.92%
Rel. target s.d.:	10.00% (Limited)	Reference value:	0.0634 mg/m ³
No. of laboratories:	28	Range of tolerance:	0.0511 - 0.0767 mg/m ³ (Z-Score <= 2.00)



Summary results

Sample:	Sample 3	Mean:	0.0922 mg/m ³
Measurand:	sulphuric acid	Reproducibility s.d.:	0.0101 mg/m ³
Method:	ISO 5725-2	Rel. reproducibility s.d.:	10.97%
Rel. target s.d.:	10.00% (Limited)	Reference value:	0.0937 mg/m ³
No. of laboratories:	29	Range of tolerance:	0.0738 - 0.1107 mg/m ³ (Z-Score <= 2.00)



Analyte chart of Z-Scores

Measurand: sulphuric acid

